



JAPANESE
INDUSTRIAL
STANDARD

Translated and Published by
Japanese Standards Association

JIS K 2536-5 : 2003
(PAJ)

**Liquid petroleum products—
Testing method of components
Part 5 : Determination of oxygenate
compounds by gas chromatography**

ICS 75.080

Reference number : JIS K 2536-5 : 2003 (E)

PROTECTED BY COPYRIGHT

 **Documents**
Document Sharing Hub

Foreword

This translation has been made based on the original Japanese Industrial Standard established by the Minister of Economy, Trade and Industry through deliberations at the Japanese Industrial Standards Committee according to the proposal of establishing a Japanese Industrial Standard from Petroleum Association of Japan (PAJ), with a draft of Industrial Standard based on the provision of Article 12 Clause 1 of the Industrial Standardization Law. Attention is drawn to the possibility that some parts of this Standard may conflict with a patent right, application for a patent after opening to the public, utility model right or application for registration of utility model after opening to the public which have technical properties. The relevant Minister and the Japanese Industrial Standards Committee are not responsible for identifying the patent right, application for a patent after opening to the public, utility model right or application for registration of utility model after opening to the public which have the said technical properties.

JIS K 2536 consists of the following 6 parts under the general title *Liquid petroleum products—Testing method of components*:

Part 1 : Fluorescent indicator adsorption method

Part 2 : Determination of total components by gas chromatography

Part 3 : Determination of aromatic components by gas chromatography

Part 4 : Determination of components by tandem type gas chromatography

Part 5 : Determination of oxygenate compounds by gas chromatography

Part 6 : Determination of oxygen content and oxygenate compounds by gas chromatography and oxygen selective detection.

Date of Establishment: 2003-06-20

Date of Public Notice in Official Gazette: 2003-06-20

Investigated by: Japanese Industrial Standards Committee
Standards Board

Technical Committee on Chemical Products

JIS K 2536-5:2003, First English edition published in 2005-01

Translated and published by: Japanese Standards Association
4-1-24, Akasaka, Minato-ku, Tokyo, 107-8440 JAPAN

In the event of any doubts arising as to the contents,
the original JIS is to be the final authority.

© JSA 2005

All rights reserved. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

Printed in Japan

PROTECTED BY COPYRIGHT

Contents

	Page
1 Scope	1
2 Normative references	2
3 Principle of test	3
4 Test instrument and device	3
4.1 Gas chromatograph	3
4.2 Data processing apparatus	4
4.3 Microsyringe	4
5 Reagents	4
6 Preparation of test	4
6.1 Adjustment of gas chromatograph.....	4
6.2 Preparation of working curve	5
7 Sampling and preparation method of sample	6
8 Procedure of test.....	6
9 Calculation method	8
10 Precision	8
11 Report of test results	8

**Liquid petroleum products—
Testing method of components
Part 5 : Determination of oxygenate
compounds by gas chromatography**

1 Scope This part of **JIS K 2536** specifies the determination of methanol and methyl-tert-butyl ether (hereafter referred to as “MTBE”) in the liquid petroleum products of which the end point is not more than 250 °C by atmospheric pressure distillation method specified in **JIS K 2254** according to gas chromatography.

- Remarks 1 The range of determination according to this Standard is 0.1 volume % to 3 volume % for methanol and 0.1 volume % to 10 volume % for MTBE.
- 2 The alcohols other than methanol may be measured by the similar method, provided that, before they are measured, the components are separated to confirm and identified.
- 3 When a doubt arises about the test results of methanol and MTBE obtained according to this Standard, the test shall be carried out according to **JIS K 2536-2**.
- 4 The use of this Standard may involve hazardous reagents, operations and equipment. This Standard does not purport to address all of the safety problems associated with its use. It is the responsibility of the user of this Standard to establish appropriate safety and health practices and determine the applicability of regulatory limitation prior to test.

Information : The test methods specified in respective parts of **JIS K 2536** are shown in informative table 1.