

JIS

JAPANESE
INDUSTRIAL
STANDARD

Translated and Published by
Japanese Standards Association

JIS T 0601-2-201 : 2005

(JIPT/JSA)

**Medical electrical equipment —
Particular requirements for the safety
of hydrotherapy and warmbaththerapy
equipment**

ICS 11.040.60

Reference number : JIS T 0601-2-201 : 2005 (E)

PROTECTED BY COPYRIGHT

 **Documents**
Document Sharing Hub

7 S

Foreword

This translation has been made based on the original Japanese Industrial Standard established by the Minister of Health, Labour and Welfare and the Minister of Economy, Trade and Industry through deliberations at the Japanese Industrial Standards Committee according to the proposal of establishing a Japanese Industrial Standard from Japan Industries Association of Physical Therapy Device (JIPT)/Japanese Standards Association (JSA), with a draft of Industrial Standard based on the provision of Article 12 Clause 1 of the Industrial Standardization Law.

Attention is drawn to the possibility that some parts of this Standard may conflict with a patent right, application for a patent after opening to the public, utility model right or application for registration of utility model after opening to the public which have technical properties. The relevant Ministers and the Japanese Industrial Standards Committee are not responsible for identifying the patent right, application for a patent after opening to the public, utility model right or application for registration of utility model after opening to the public which have the said technical properties.

Date of Establishment: 2005-03-25

Date of Public Notice in Official Gazette: 2005-03-25

Investigated by: Japanese Industrial Standards Committee

Standards Board

Technical Committee on Medical Equipment

JIS T 0601-2-201 : 2005, First English edition published in 2005-10

Translated and published by: Japanese Standards Association
4-1-24, Akasaka, Minato-ku, Tokyo, 107-8440 JAPAN

In the event of any doubts arising as to the contents,
the original JIS is to be the final authority.

© JSA 2005

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

Printed in Japan

PROTECTED BY COPYRIGHT

MT
Documents
Document Sharing Hub

Contents

	Page
Introduction	1
Section 1 General	1
1 Scope and object.....	1
2 Definitions	2
6 Identification, marking and documents.....	3
Section 2 Environmental conditions.....	3
Section 3 Protection against electric shock hazards.....	3
13 General	3
Section 4 Protection against mechanical hazards	4
27 Pneumatic and hydraulic power	4
Section 5 Protection against hazards from unwanted or excessive radiation	4
36 Electromagnetic compatibility	4
Section 6 Protection against hazards of ignition of flammable anesthetic mixtures.....	4
Section 7 Protection against excessive temperatures and other safety hazards	4
Section 8 Accuracy of operating data and protection against hazardous output	4
50 Accuracy of operating data	4
51 Protection against hazardous output.....	5
Section 9 Abnormal operation and fault conditions : environmental tests	6
Section 10 Constructional requirements	6
Annex AA (normative) Method for measuring ultrasound wave output generated from air bubbles.....	7

Medical electrical equipment— Particular requirements for the safety of hydrotherapy and warmbaththerapy equipment

Introduction This Japanese Industrial Standard shall be read along with the general standard JIS T 0601-1 (*Medical electrical equipment—Part 1: General requirements for safety*) because this Standard is composed by referring to JIS T 0601-1.

The numbering of sections, clauses and subclauses of this Standard corresponds with that of JIS T 0601-1. The changes to the text of JIS T 0601-1 are specified by the use of the following “words”:

“**Replacement**” means that the clause or subclause of JIS T 0601-1 is replaced completely by the text of this Standard.

“**Addition**” means that the text of this Standard is additional to the requirements of JIS T 0601-1.

“**Amendment**” means that the clause or subclause of JIS T 0601-1 is amended as indicated by the text of this Standard.

Subclauses or figures which are additional to those of JIS T 0601-1 are numbered starting from 101, additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc. Terms defined in clause 2 of this Standard or of JIS T 0601-1 are boldfaced.

Section 1 General

Section 1 of JIS T 0601-1 is applied here with the following changes.

1 Scope and object

1.1 Scope

Addition

This Standard specifies requirements for the safety of **hydrotherapy and warmbaththerapy equipment** used by practitioners for physical therapy, which is defined in 2.1.101 and 2.1.102.

This Standard does not apply to:

- vapor pressure **equipment**,
- household **equipment**,
- **equipment** whose performance cannot be tested unless they are assembled at installation location,
- **equipment** which is used with warm water mixed with radioactive ingredients.