

ERRATA

(continued to the next page)

Page 17

Figure 12

Replace the table in the figure with the following:

Maximum user mass	Force applied to push handle	
	Force applied to either left or right handle, in the case of two separate handles (F)	Force applied to either side of the handle, in the case of one unseparated handle ($F \times 2$)
Up to and incl. 25 kg	330 N $\pm$ 10 N	660 N $\pm$ 20 N
Over 25 kg up to and incl. 50 kg	520 N $\pm$ 16 N	1 040 N $\pm$ 32 N
Over 50 kg up to and incl. 75 kg	700 N $\pm$ 21 N	1 400 N $\pm$ 42 N
Over 75 kg up to and incl. 100 kg	880 N $\pm$ 26 N	1 760 N $\pm$ 52 N

Page 22

10.2.14

Error: The composition of test apparatus shall be as specified in Annex 5, and the preparation of test apparatus and the fitting method of dummy be as specified in **ISO 7176-8**.

Correct: The composition of test apparatus shall be as specified in Annex 5, and the preparation of test apparatus, the fitting method of dummy and other matters shall be as specified in **ISO 7176-8**.

Annex 3 Table 1

Replace the table with the following:

Annex 3 Table 1 Classification of dummies

Applicable wheelchair types	For 100 kg	For 75 kg	For 50 kg	For 25 kg
Nominal mass of corresponding dummy	100 kg	75 kg	50 kg	25 kg
Trunk	9 weights of 6 kg 1 weight of 3 kg	7 weights of 6 kg	4 weights of 6 kg	2 weights of 6 kg
	4 kg ⁽¹⁾	4 kg ⁽¹⁾	4 kg ⁽¹⁾	1.5 kg ⁽¹⁾
Subtotal	61 kg ± 3 kg	46 kg ± 3 kg	28 kg ± 3 kg	13.5 kg ± 2 kg
Thigh	4 weights of 6 kg 1 weight of 3 kg	3 weights of 6 kg	2 weights of 6 kg	1 weight of 6 kg 1 weight of 3 kg
	4 kg ⁽¹⁾	4 kg ⁽¹⁾	4 kg ⁽¹⁾	1.5 kg ⁽¹⁾
Subtotal	31 kg ± 3 kg	22 kg ± 3 kg	16 kg ± 3 kg	10.5 kg ± 2 kg
Leg (combined)	1 weight of 6 kg	1 weight of 6 kg	1 weight of 6 kg	
	1 kg ⁽¹⁾	1 kg ⁽¹⁾	1 kg ⁽¹⁾	
Subtotal	7 kg ± 1 kg	7 kg ± 1 kg	7 kg ± 1 kg	
Leg (separate)	2 weights of 3.5 kg ± 0.5 kg, of size 75 mm × 150 mm × 40 mm will be appropriate.	2 weights of 3.5 kg ± 0.5 kg, of size 75 mm × 150 mm × 40 mm will be appropriate.	2 weights of 3.5 kg ± 0.5 kg, of size 75 mm × 150 mm × 40 mm will be appropriate.	
Subtotal	7 kg ± 1 kg	7 kg ± 1 kg	7 kg ± 1 kg	
Total mass	100 ⁺⁵ ₋₂ kg	75 ⁺⁵ ₋₂ kg	50 ⁺⁵ ₋₂ kg	25 ⁺⁵ ₋₂ kg

Note (1): Mass of every part of the dummy other than weights.

If this mass is not 4 kg, regardless of the material, the number of the weights and the fixing position shall be adjusted and the centre position of mass shall be as shown in Annex 3 figure 2.

Remarks: This erratum is for correcting the first edition of this Standard.

Japanese Standards Association